

BLANK PAGE






Indian Standard

SPECIFICATION FOR SUCTION TUBES WITH STYLET MICROLARYNGEAL SURGERY

- 1. Scope Covers dimensional and other requirements for suction tube with stylet used in microlaryngeal ENT surgery.
- 2. Shape and Dimensions As shown in Fig. 1.




All dimensions in millimetres.

FIG. 1 SUCTION TUBE WITH STYLET MICROLARYNGEAL SURGERY

- 2.1 A deviation of $\pm 2^{\circ}5$ percent shall be allowed on all dimensions.
- 2.2 A deviation of $\pm 2.5^{\circ}$ shall be allowed on all angles.
- 3. Material Stainless steel conforming to Designation 04Cr18Ni10 or 07Cr18Ni19 of Schedule V of IS: 1570 (Part V)-1972 'Schedules for wrought steels: Part V Stainless and heat resisting steels (first revision)' or brass conforming to Designation CuZn37 or CuZn40 of IS: 410-1967 'Specification for rolled brass plate, sheet, strip and foil (second revision)'.
- 4. Workmanship and Finish
- 4.1 The suction tube shall be free from scales, pits, burrs, cracks and other surface defects.
- 4.2 The suction tube shall be well formed.
- 4.3 Welding, soldering or brazing shall be neat and sound.
- 4.4 The suction tube shall be dull matt finished and passivated.

Adopted 28 June 1978

© October 1978, ISI



IS: 8837 - 1978

- 4.5 The suction tube shall be accompanied with the proper stylet.
- 4.6 The suction tube if made of brass shall be plated nickel and chromium conforming to Grade 2 of IS: 4827-1968 'Specification for electroplated coatings of nickel and chromium on copper and copper alloys'.

5. Test

- 5.1 The suction tube if made of stainless steel shall be tested as given in 5.1.1.
- **5.1.1** Corrosion resistance Scrub the sample with soap and warm water. Rinse in hot water and then dip in 90 percent ethyl alcohol. Dry the sample. Immerse in copper sulphate solution at room temperature for 5 minutes and wash off with fresh water or wet cotton wool.
 - 5.1.2 The copper sulphate solution shall be prepared as follows:

Copper sulphate (CuSO ₄ .5H ₂ O)	4·0 g
Sulphuric acid (H ₂ SO ₄) (sp gr 1 [.] 84)	10 [.] 0 g
Distilled water [see IS: 1070-1977	90.0 ml
Specification for water for general	
laboratory use (second revision)]	

- 5.1.3 No red stains or spots in the sample shall be allowed.
- 6. Marking Each suction tube shall be marked with manufacturer's name, initials or recognized trade-mark, size, country of manufacture and the words 'SS' if made of stainless steel.
- 6.1 ISI Certification Marking Details available with the Indian Standards Institution.
- 7. Packing Each suction tube shall be wrapped in a moisture-proof paper and put in a cardboard or polyethylene bag. The packing may also be done as agreed to between the purchaser and the supplier.